



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Geoform, Inc.
16832 Gramercy Place
Gardena, CA 90247

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 21 November 2025
Certificate Number: AD-2822



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Geoform, Inc.
16832 Gramercy Place
Gardena, CA 90247
Steve Farentinos 424-292-3407

DIMENSIONAL MEASUREMENT

Valid to: **November 21, 2025**

Certificate Number: **AD-2822**

1 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 1D	Up to 1 in	140 μ m	Outside Micrometers utilized as Reference Standard for Dimensional Measurement.
	Up to 6 in Up to 12 in	1 400 μ m 2 600 μ m	Calipers utilized as Reference Standard for Dimensional Measurement.

3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 3D ¹	X = Up to 1 000 mm Y = Up to 1 200 mm Z = Up to 600 mm	(4.3 + 0.014L) μ m	Coordinate Measuring Machine utilized as Reference Standard for Dimensional Measurement.
	X = Up to 1 000 mm Y = Up to 1 600 mm Z = Up to 600 mm	(6.3 + 0.012L) μ m	Coordinate Measuring Machine utilized as Reference Standard for Dimensional Measurement.
	X = Up to 900 mm Y = Up to 1 200 mm Z = Up to 800 mm	(4.5 + 0.013L) μ m	Coordinate Measuring Machine utilized as Reference Standard for Dimensional Measurement.


3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 3D ¹	Up to 2.5 m	$(77 + 0.014L) \mu\text{m}$	Articulating Arm Coordinate Measuring Machine utilized as Reference Standard for Dimensional Measurement.
	Up to 2.5 m	$(89 + 0.021L) \mu\text{m}$	Articulating Arm Coordinate Measuring Machine with Scanner (Laser Line Probe) utilized as Reference Standard for Dimensional Measurement.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. L = length in mm.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AD-2822.



Jason Stine, Vice President